
TWL Calibration Data Summary

Serial Number: 0301359

Full-Scale Flow Rate: 80
(yields 20mA output)

Zero Flow Rate: (yields 4mA output)

Bridge Resistance: 350.2 ohms

Full-Scale Bridge Output: 2.0987 mV/V

Calibration Resistor (R7): 47205 ohms

Calibration Voltage: 0.8805 volts

Calibration Flow Rate: 75.06

Calibration Current: 19.01 mA

Zero Switch Setting: 0

TB3 Jumper Position: + (POSITIVE)

Span Switch Setting - Undamped: 0

Span Switch Setting - Damped: 8

Additional Information:
